

Conforms to JIS A Class and DIN Class C

CRI Illuminance Meter



CAL

INOLTA

Easy to use !

Can take spectral measurements of flash light

Good tool for lighting planners

Rotating receptor head



Dark calibration

performed

needing cap

without



Easy-to-read display *Display mode examples



Color rendering index



CIE1931 (CIE1964)



584lx 4397K 0.0018 0 3653 0.3706 Text



584lx

1397







Measurement and evaluation of the illuminance, color-temperature, and colorrendering index of indoor illumination sources such as LEDs, fluorescent lamps, etc.





Giving Shape to Ideas

System Diagram



Main Specifications of CL-70F

Model	CRI Illuminance Meter CL-70F
Illuminance meter class	Conforms to requirements for Class A of JIS C1609-1 : 2006 "Illuminance meters Part1:General measuring instruments; Conforms to DIN 5032 Part 7 Class C
Sensor	CMOS linear image sensor
Spectral wavelength range	380 nm to 780 nm
Output wavelength pitch	1 nm
Measuring range	Constant light: 1 to 200,000 lx; 1,563 to 100,000 K (Chromaticity display requires 5 lx or more) Flash light: 20 to 20,500 lx • s; 2,500 to 100,000 K
Accuracy (Standard Illuminant A) (*1, 2)	E_{v} : ±5%±1digit of displayed value
	xy: ±0.003 (at 800 lx)
Repeatability (Standard Illuminant A) (*1)	$E_{\nu}\!\!:$ 30 to 200,000 lx: 1%+1digit; 1 to 30 lx: 5%+1digit (*3)
	xy: 500 to 200,000 lx: 0.001 (*4) xy: 100 to 500 lx: 0.002 (*4) xy: 30 to 100 lx: 0.004 (*4) xy: 5 to 30 lx: 0.008 (*4)
Visible-region relative spectral response characteristics (f1')	Within 9%
Cosine correction characteristics (f2)	Within 6%
Temperature drift (f_T)	E _v : ±5% xy: ±0.006
Humidity drift (f _H)	E _v : ±3% xy: ±0.006
Power	2 AA-size batteries (Alkaline batteries or manganese dry cells); USB bus power
Response time	Constant light (Maximum): 15 sec Constant light (Minimum): 0.5 sec Flash light: 1 to 1/500 sec (in 1-step intervals) (*5)
Color indication modes	Correlated color temperature T _{cp} , Difference from blackbody Δ uv, XYZ, xy, u'v', Dominant wavelength λ_d , Excitation purity P _e , Spectral irradiance, E _v , CRI (Ra, Ri), Peak wavelength λ_p , exposure value
Other functions	Data memory: 999 data; Preset function; Auto power off function
Display languages	English, Japanese, Chinese (Simplified)
Interface	USB 2.0 Mini B
Operation temperature/ humidity range	-10 to 40°C , relative humidity of 85% or less (at 35°C) with no condensation
Storage temperature/ humidity range	-10 to $45^\circ \! C$, relative humidity of 85% or less (at 35 $^\circ \! C$) with no condensation
Size	$73 (W) \times 183 (H) \times 27 (D) mm$ (Not including projecting buttons) D (max): 40 mm
Weight (without battery)	230 g

(*1) Measurement mode: Constant light (range L), Exposure time : AUTO

(*2) Linear for Ev

(*3) 10 times measurement (2σ) /Ave
(*4) 10 times measurement (2σ)

(*5) Shutter speed

Dimensions (Units: mm)



Utility Software (Standard accessory)

Software	OS
Windows	Windows [®] 10 Pro 32bit/64bit Windows [®] 11 Pro

· Windows[®] is trademarks of Microsoft Corporation in the USA and other countries.

Konica Minolta's Illuminance Measurement Trio <All conforms to JIS AA Class>

Illuminance Spectrophotometer CL-500A For industrial applications requiring high accuracy





Chroma Meter CL-200A A de facto industry standard for colortemperature measurement !



Illuminance Meter T-10A Capable of accurately measuring next-generation lamps including PWM-controlled lighting.



* Both CL-200A and CL-500A also can measure PWM-controlled light.

KONICA MINOLTA, the Konica Minolta logo and symbol mark, and "Giving Shape to ideas" are registered trademarks or trademarks of KONICA MINOLTA, INC.

The specifications and appearance shown herein are subject to change without notice. Screens shown are for illustration purpose only.

• Some lamp control methods may make accurate measurements difficult. For details, please contact your nearest Konica Minolta sales office or dealers.



SAFETY PRECAUTIONS For correct use and for your safety, be sure to read the instruction manual before using the

instrument. Be sure to use the specified batteries. Using improper batteries may cause a fire or electric



CONTACT US - Global Network

https://www.konicaminolta.com/ instruments/network/index.html



shock