

Display Color Analyzer

Support for LED backlights

The next-generation model that surpasses the CA-210 For high-speed, high-accuracy measurements






Giving Shape to Ideas

Enables high-accuracy adjustment of EL/LED-backlit LCD TV gamma/white balance to greatly improve efficiency.

White balance adjustment has advanced even further!

Our previous Display Color Analyzer CA-210 could adjust the white balance of LED-backlit LCD TVs to $\Delta xy=0.010$, but the new Display Color Analyzer CA-310 enables adjustment to $\Delta xy=0.003$ so colors are even more true, as can be seen below.

White balance adjustment of LED -backlit LCD TVs




Enables high-speed measurement of even extremely low luminances down to 0.005 cd/m²

Sensor noise reduction technology has been used to enable measurements even in the extremely low luminance region around 0.005 cd/m² at speeds as fast as 4 times per second. This allows the high-speed high-accuracy measurement essential for manufacturing high-grade displays. In addition, at luminances higher than 2.0 cd/m², 20 measurements per second are possible.




Reduces errors due to LED emission distribution variations to less than 1/3.

Variations in the emission distribution of LED backlights result in individual differences of about 10nm in peak intensity wavelength. If LED-backlit LCD TVs with such individual differences are adjusted using conventional color analyzers, color differences of close to 0.010 on the xy chromaticity diagram may occur. But the CA-310 has sensor sensitivities that more closely match the CIE 1931 color-matching functions, enabling the color difference in the same case to be reduced to around 0.003, suppressing errors to less than 1/3.



Wavelength (nm)




*Errors (differences from true values) for white LEDs with different peak wavelengths when measured using CA-310. User calibration to standard LED performed.

Measurement errors for LED backlights

PC Software for Color Analyzer **CA-SDK (Standard accessory)**

Standard accessory SDK helps create software easily according to needs.

Sample software is bundled; you can start data collection easily.



Example of White Balance Adjustment Software made by SDK

Required system


OS: Windows® XP, Vista, 7

board CA-B15)

Number of digits for luminance

Windows® and Excel® are a trademark of Microsoft Corporation in the USA and other countries.

probes. (Requires expansion




Sample software Gamma

Sample software (Standard)

- Cal CA-210 can be corrected in the matrix calibration method using Konica Minolta's spectroradiometer CS-1000A. Color The measurement data of CA-210 can be acquired into the PC. Drift tests, LCD stability test and so on can be performed easily. The acquired data can be read with Excel® or other spreadsheet software. Contrast Multi-point measurement (5, 9, or 25 points) can be made for
- white uniformity and contrast measurement.
- R, G, B, and W gamma measurements for gradations of 16, 32, Gamma 64, 128, and 256 steps.

CA-310 Probe



Probe variations

0.0001 cd/m².

This table is based on the most popular method for controlling emission intensity for each display type.

* Measurements of displays using certain control methods are not possible. For details of measurement compatibility, contact your nearest Konica Minolta representative.

Examples for which measurement is not possible:

- Displays which use PWM, etc. for control of emission intensity.
- Displays with backlights which emit intermittently.
- Displays which write black for each frame

etc.


- \cap Recommended
- Measurement \triangle possible with restrictions, but probes marked with O are recommended
- × Measurement not possible


e black for each frame,			LED Universal Measuring Probe		LED Flicker Measuring Probe	
			Ø27 Probe CA-PU32 (2m) CA-PU35 (5m)	Ø10 Probe CA-PSU32 (2m) CA-PSU35 (5m)	Ø27 Probe CA-P32 (2m) CA-P35 (5m)	Ø10 Probe CA-PS32 (2m) CA-PS35 (5m)
Applicability for different display types						
Transmissive / semi-transmissive LCD		Active Matrix Driven	0	0	0*	0*
		Passive Matrix Driven	0	0	×	×
OLED		Active Matrix Driven	0	0	0*	0*
		Passive Matrix Driven	0	0	×	×
PDP			0	Δ	×	×
FED		0	0	×	×	
Rear Screen Projector	LCD	Active Matrix Driven	0	\triangle	0*	\triangle^*
		Passive Matrix Driven	0	Δ	×	×
	DLP		0	Δ	×	×
	CRT		0	Δ	×	×
LED Flicker Measuring Pr	obes are	unsuitable for measurements of (CRTs.)			

System Diagram

Specifications

Model		CA-310(LED Universal Measuring Ø27 Probe)	CA-310 (LED Universal Measuring Ø10 Probe)	CA-310 (LED Flicker Measuring Ø27 Probe)	CA-310 (LED Flicker Measuring Ø10 Prob				
Detector		Silicon photo cell	Cristo (LED Oniversative asuning Dio Flobe)	Terroro (EED Flicker Measuring 027 Flobe)	Terroro (LED Fricker Measuring D10 F1000				
Measurement area		Ø27mm Ø10 mm		Ø27 mm	Ø10 mm				
Acceptance angle		±2.5°	±5°	±2.5°	±5°				
Measurement distance		30±10 mm	30±5 mm	30±10 mm	30±5 mm				
Display	Luminance	0.0001 to 1000 cd/m ²	0.0001 to 3000 cd/m ²	0.0001 to 1000 cd/m ²	0.0001 to 3000 cd/m ²				
range	Chromaticity	Displayed in 4 or 3-digit value (Can be chose							
Luminance	Measurement range		0.0150 to 3000 cd/m ²	0.0050 to 1000 cd/m ²	0.0150 to 3000 cd/m ²				
	Accuracy	0.0050 to 0.0999 cd/m ² ±4%±0.0015 cd/m ²	0.0150 to 0.2999 cd/m ² ±4%±0.0045 cd/m ²	0.0050 to 0.0999 cd/m ² ±4%±0.0015 cd/m ²	0.0150 to 0.2999 cd/m ² ±4%±0.0045 cd/r				
	(for white)*1	0.1000 to 9.999 cd/m ² ±3%±0.0010 cd/m ²	0.3000 to 29.99 cd/m ² ±3%±0.0030 cd/m ²	0.1000 to 9.999 cd/m ² ±3%±0.0010 cd/m ²	0.3000 to 29.99 cd/m ² ±3%±0.0030 cd/r				
		10.00 to 1000 cd/m ² ±2%±0.0010 cd/m ²	30.00 to 3000 cd/m ² ±2%±0.0030 cd/m ²	10.00 to 1000 cd/m ² ±2%±0.0010 cd/m ²	30.00 to 3000 cd/m ² ±2%±0.0030 cd/r				
	Repeatability(20) *1	0.0050 to 0.0999 cd/m ² 1% + 0.0010 cd/m ²	0.0150 to 0.2999 cd/m ² 1% + 0.0030 cd/m ²	0.0050 to 0.0999 cd/m ² 1% + 0.0010 cd/m ²	0.0150 to 0.2999 cd/m ² 1% + 0.0030 cd/m				
		0.1000 to 0.9999 cd/m ² 0.2% + 0.0010 cd/m ²	0.3000 to 2.999 cd/m ² 0.2% + 0.0030 cd/m ²	0.1000 to 0.9999 cd/m ² 0.2% + 0.0010 cd/m ²	0.3000 to 2.999 cd/m ² 0.2% + 0.0030 cd/				
		1.000 to 1000 cd/m ² 0.1%+0.0010 cd/m ²	3.000 to 3000 cd/m ² 0.1% + 0.0030 cd/m ²	1.000 to 1000 cd/m ² 0.1%+0.0010 cd/m ²	3.000 to 3000 cd/m ² 0.1% + 0.0030 cd/				
Chromatcity	Measurement range	0.0500 to 1000 cd/m ²	0.1500 to 3000 cd/m ²	0.0500 to 1000 cd/m ²	0.1500 to 3000 cd/m ²				
	Accuracy *1	0.0500 to 4.999 cd/m ² ±0.005 for white	0.1500 to 14.99 cd/m ² ±0.005 for white	0.0500 to 4.999 cd/m ² ±0.005 for white	0.1500 to 14.99 cd/m ² ±0.005 for whi				
	(temperature:23°±2°,	5.000 to 19.99 cd/m ² ±0.004 for white	15.00 to 59.99 cd/m ² ±0.004 for white	5.000 to 19.99 cd/m ² ±0.004 for white	15.00 to 59.99 cd/m ² ±0.004 for whi				
	relative humidity:	20.00 to 1000 cd/m ² ±0.003 for white	60.00 to 3000 cd/m ² ±0.003 for white	20.00 to 1000 cd/m ² ±0.003 for white	60.00 to 3000 cd/m ² ±0.003 for whi				
	(40±10)%))	120 cd/m ² ±0.002 for white	120 cd/m ² ±0.002 for white	120 cd/m ² ±0.002 for white	120 cd/m ² ±0.002 for whi				
		(±0.004 for monochrome)*2	(±0.004 for monochrome)*2	(±0.004 for monochrome)*2	(±0.004 for monochrome)				
	Repeatability(2o) *1	0.0500 to 0.0999 cd/m ² 0.010	0.1500 to 0.2999 cd/m ² 0.010	0.0500 to 0.0999 cd/m ² 0.010	0.1500 to 0.2999 cd/m ² 0.010				
		0.1000 to 0.1999 cd/m ² 0.004	0.3000 to 0.5999 cd/m ² 0.004	0.1000 to 0.1999 cd/m ² 0.004	0.3000 to 0.5999 cd/m ² 0.004				
		0.2000 to 0.4999 cd/m ² 0.002	0.6000 to 1.499 cd/m ² 0.002	0.2000 to 0.4999 cd/m ² 0.002	0.6000 to 1.499 cd/m ² 0.002				
		0.5000 to 1000 cd/m ² 0.001	1.500 to 3000 cd/m ² 0.001	0.5000 to 1000 cd/m ² 0.001	1.500 to 3000 cd/m ² 0.001				
Flicker	Measurement range		-	5 cd/m ² or higher	15 cd/m² or higher				
Contrast	Display range		-	0.0 ~ 999.9 %					
method	Accuracy		-	±1 % (Flicker frequency: 30 Hz AC/DC 10%	sine wave)				
				±2 % (Flicker frequency: 60 Hz AC/DC 10% sine wave)					
	Repeatability(2o)		-	1 % (Flicker frequency: 20 to 65 Hz AC/DC 10% sine wave)					
Flicker JEITA			-	5 cd/m² or higher	15 cd/m ² or higher				
method *3	Accuracy		-	± 0.5 dB (Flicker frequency: 30 Hz AC/DC 4% (-40 dB) sine wave) ± 1.0 dB (Flicker frequency: 30 Hz AC/DC 1.2% (-50 dB) sine wave)					
	Repeatability(2o)		-	0.1 dB (Flicker frequency: 30 Hz AC/DC 4% (-40 dB) sine wave) 0.3 dB (Flicker frequency: 30 Hz AC/DC 1.2% (-50 dB) sine wave)					
Measure-	xyL _v	0.0050 to 0.0999 cd/m ² 4(3.5) times/sec.	0.0150 to 0.2999 cd/m ² 4(3.5) times/sec.	0.0050 to 0.0999 cd/m ² 4(3.5) times/sec.	0.0150 to 0.2999 cd/m ² 4(3.5) times/set				
ment		0.1000 to 1.999 cd/m ² 5(4.5) times/sec.	0.3000 to 5.999 cd/m ² 5(4.5) times/sec.	0.1000 to 1.999 cd/m ² 5(4.5) times/sec.	0.3000 to 5.999 cd/m ² 5(4.5) times/se				
speed*4		2.000 to 1000 cd/m ² 20(17) times/sec.	6.000 to 3000 cd/m ² 20(17) times/sec.	2.000 to 1000 cd/m ² 20(17) times/sec.	6.000 to 3000 cd/m ² 20(17) times/se				
	Flicker Contrast		-	16(16) times/sec.					
	Flicker JEITA *3		-	0.5 (0.3)times/sec.*5					
Display	Digital	xyL _v , T⊿uvL _v , RGB analyze, XYZ, u'v'L _v		xyL _v , T <u>/</u> uvL _v , RGB analyze, XYZ, u'v'L _v , Flicker (Contrast method) *3					
	Analog	$\Delta x \Delta y \Delta L_v$, R/G B/G Δ G, Δ R B/R G/R		∆x∠y∠L _v , R/G B/G ∠G, ∠R B/R G/R, Flicker (Contrast method) *3					
	LCD	16 characters by 2 lines (with backlight)							
SYNC mode		NTSC, PAL, EXT, UNIV, INT							
Object under measurement		Vertical synchronization frequency: 40 to 2	200 Hz	Vertical synchronization frequency: 40 to 200 Hz (Luminance or chromaticity measurement), 40 to 130 Hz (Flicker measurement)					
Memory channel		100 channels							
Analyzer function		Standard function							
Interface		USB; RS-232C (38,400 bps or below)							
Multi-point	Measurement	Max. 5 points (Use 4-Probe Expansion Boar	rd CA-B15)						
Operation tem	perature/humidity range	Temperature: 10 to 28°C; relative humidity Chromaticity change ±0.002 for white, ±0.0	70% or less with no condensation Lumi 006 for monochrome from reading of Konic	nance change: ±2% of reading for white a Minolta's standard LCD *1, 120 cd/m², with	23°C 40%				
Storage tempe	erature/humidity range		h no condensation 28 to 40°C : relative hum						
Input voltag	e range	100-240V∿, 50-60 Hz, 50 VA		,					
	Main body	340(W)×127(H)×216(D) mm/3.58 kg							
	Probe	Ø49×208 mm / 530 g	Ø49×236 mm / 550 g	Ø49×208 mm / 530 g	Ø49×236 mm / 550 g				





• Select the desired type of LED Universal Measuring

236

47.5

Phone : +31 (0) 30 248-1193 Phone : +49 (0) 89 4357 156 0 Phone : +33 (0) 1 80 11 10 70

Phone : +43(0) 1926 267300 Phone : +44(0) 1926 267300 Phone : +41(0) 43 322-9800 Phone : +46(0) 31 7099464 Phone : +46(0) 71 73452-11 Phone : +90(0) 216-528 56 56 Phone : +86-(0) 21-5489 0202 Phone : +86-(0) 21-5489 0202

Phone: 888-473-2656 (in USA), 201-236-4300 (outside USA)

Nieuwegein, Netherlands München, Germany Roissy CDG, France Warrington, United Kingdom Cinisello Balsamo, Italy Dictikop Switzerdoad

Dietikon, Switzerland Västra Frölunda, Sweden

Wroclaw, Poland Istanbul, Turkey Shanghai, China

Beijing, China Guangdong, China Chongqing, China Shandong, China

Hubei, China

Singapore Goyang-si, Korea

- Contains mercury in the backlighting of LCD used for display. Dispose of according to local, state or federal laws.
- KONICA MINOLTA, the Konica Minolta logo and symbol mark, and "Giving Shape to ideas" are registered
- trademarks or trademarks of KONICA MINOLTA, INC. Screens shown are for illustration purpose only. • The specifications and appearance shown herein are
- subject to change without notice. Some lighting control methods may make accurate
- measurements difficult. For details, please contact your nearest Konica Minolta sales office or dealer.

Fax: 201-785-2482

Fax : +31(0), $302 \pm 481 \pm 211$ Fax : +49(0), 89435715699Fax : +33(0), 180111082Fax : +40(0), 1925711143Fax : +4902849488.30Fax : +41(0), 43322-9809



SAFETY PRECAUTIONS For correct use and for your safety, be sure to read the instruction manual before using the instrument. • Always connect the instrument to the specified power supply voltage. Improper connection may cause a fire or electric shock



Certificate No.: JQA-QMA15888 Registration Date : October 26, 2018 Registration Date : October 26, 2018 KONICA MINOLTA, Inc., Sakai Site oduct design, manufacture/manufacturing management, calibration, and service



Certificate No : JQA-E-80027 Registration Date : March 12, 1997 KONICA MINOLTA, Inc., Sakai Site

KONICA MINOLTA, INC. Konica Minolta Sensing Americas, Inc. Konica Minolta Sensing Europe B.V.	Osaka, Japan New Jersey, U.S.A. European Headquarter /BENELUX German Office French Office UK Office Italian Office Swiss Office Nordic Office Polish Office Turkish Office
Konica Minolta (CHINA) Investment Ltd.	SE Sales Division
	Beijing Office
	Guangzhou Office
	Chongqing Office
	Qingdao Office
	Wuhan Office
Koniga Minalta Canaing Cinganara Bta Ltd	

80

108

Konica Minolta Sensing Singapore Pte Ltd. Konica Minolta Sensing Korea Co., Ltd.

Addresses and telephone/fax numbers are subject to change without notice. For the latest contact information, please refer to the KONICA MINOLTA Worldwide Offices web page :

©2013 KONICA MINOLTA, INC.

Fax : +48 (0)71 734 52 10 Fax : +90 (0) 212-253 49 69 Fax : +86- (0)21-5489 0005 Fax : +86- (0)20-8522 1241 Fax : +86- (0)20-3826 4223 Fax : +86- (0)20-3826 4223 Phone : +86- (0)21-8463 6262 Phone : +86- (0)20-3826 4220 Phone : +86- (0)23-6773 4988 Phone: +86-(0)532-8079 1871 Fax: +86- (0)532-8079 1873 Phone : +86 (0)27-8544 9942 Phone : +65 6563-5533 **Fax**: +86- (0)27-8544 9991 **Fax**: +65 6560-9721 Phone: +82(0)2-523-9726 Fax: +82(0)31-995-6511

https://konicaminolta.com/instruments/network

9242-4885-16 BIMPK Printed in Japan